

Calibration Certificate

This is to state that the Item below has been fully tested and calibrated as follows:

Product: Xograph "Digit-X" Densitometer

Serial No: L30468

Calibration Standard: Transmission Density Stepwedge S/N: 250308
As Described Under ASME SE-1079 & ASTM E 1079-00

Light Source: View-Lite 0417L LED film viewer
At 100,000 cd/m²

Tested & Calibrated By: Signature: C. Morris

Test & Calibration Date: 19/02/26

Recalibration Date: 19/02/27

Calibration Results*

Step Number	1	3	5	8	10	11	12	13	14
Actual Density	0.17	0.67	1.26	2.17	2.84	3.24	3.54	3.91	4.45
Measured Density	0.19	0.66	1.24	2.16	2.83	3.21	3.51	3.89	4.46

Reference XHC338 V4.0

Xograph Healthcare Limited, Xograph House, Ebley Road, Stonehouse, Gloucestershire GL10 2LU
T +44 (0) 1453 820 320 E enquiry@xograph.com
xograph.com

Directors: Nathan Pettitt (CEO), Stephen Arundel, Darragh Dreenan
Company Reg. No. 00918971